

# EC CERTIFICATE

Number: 86443CE09

## Production Quality Assurance

**Directive 93/42/EEC on Medical devices, Annex V**  
(Devices in Class IIa, IIb or III)

Manufacturer:

**Philips Consumer Lifestyle B.V. (NL18/805060)**

**Tussendiepen 4  
9206 AD Drachten  
The Netherlands**

For the product category(ies)

### Digital thermometer

DEKRA grants the right to use the EC Notified Body Identification Number illustrated below to accompany the CE Marking of Conformity on the products concerned conforming to the required Technical Documentation and meeting the provisions of the EC-Directive which apply to them:

# 0344

Documents, that form the basis of this certificate:

**Certification Notice 86443CN, initially dated 29 June 1998**  
**Addendum, initially dated 13 January 2014**

DEKRA hereby declares that the above mentioned manufacturer fulfils the relevant provisions of 'Besluit Medische Hulpmiddelen', the Dutch transposition of the Council Directive 93/42/EEC of June 14, 1993 concerning Medical devices, including all subsequent amendments. The manufacturer has implemented a quality assurance system for the manufacture and final inspection for the above mentioned product category in accordance to the provisions of Annex V of Council Directive 93/42/EEC of June 14, 1993 and is subject to periodical surveillance. For placing on the market of Class III or Class IIb devices an additional EC type-examination certificate according to Annex III is mandatory. The necessary information related to the quality assurance system of the manufacturer, including facilities and the reference to the relevant documentation, of the products concerned and the assessments performed, are stated in the Certification Notice which forms an integrative part of this certificate.

This certificate is valid until: 15 May 2018  
Issued for the first time: 13 January 2014  
Reissued: 15 May 2015

DEKRA Certification B.V.



drs. G.J. Zoetbrood  
Managing Director



ing. A.A.M. Laan  
Certification Manager

© Integral publication of this certificate and adjoining reports is allowed

DEKRA Certification B.V. is Notified Body with ID no 0344

DEKRA Certification B.V. Meander 1051, 6825 MJ Arnhem P.O. Box 5185, 6802 ED Arnhem, The Netherlands  
T +31 88 96 83000 F +31 88 96 83100 [www.dekra-certification.com](http://www.dekra-certification.com) Company registration 09085396



# ADDENDUM

Belonging to certificate: 86443CE09

1/1

## CE MARKING OF CONFORMITY MEDICAL DEVICES

Digital thermometer

Issued to:

**Philips Consumer Lifestyle B.V. (NL18/805060)**  
Tussendiepen 4  
9206 AD Drachten  
The Netherlands

This certificate covers the following product(s):

- Philips AVENT digital thermometer SCH410
- Philips ear thermometer SCH740
- Philips ear thermometer DL8740

Initial date: 13 January 2014

Revision date: 2 December 2015

DEKRA Certification B.V.



drs. G.J. Zoetbrood  
Managing Director



ing. A.A.M. Laan  
Certification Manager

© Integral publication of this certificate and adjoining reports is allowed

DEKRA Certification B.V. is Notified Body with ID no 0344

DEKRA Certification B.V. Meander 1051, 6825 MJ Arnhem P.O. Box 5185, 6802 ED Arnhem, The Netherlands  
T +31 88 96 83000 F +31 88 96 83100 [www.dekra-certification.com](http://www.dekra-certification.com) Company registration 09085396